



3902 W. Valley Hwy. N. Bldg 502,
Auburn, Wa. 98001 USA
253.735.3910

DATE OF ISSUE: 04122005

PERPHORMAX™

INSTALLATION INSTRUCTIONS



IVD

EC

REP

CE Partner4U, Esdoomlaan 13,
3951DB Maam, The Netherlands

Cartridge, Reference Electrode

PN: 373039 HI

Type: Solid state - silver/silver chloride matrix.

Use: This electrode is for *in vitro* diagnostics use only. Used as a reference electrode on Hitachi 900 series, and Modular Analytics chemistry analyzers.

Precautions: This electrode has been tested with candidate measurement electrodes (NA⁺ K⁺ CL⁻) for control recoveries using Beckman Decision, BioRad Lypocheck serum/urine, Roche Precinorm/Precipath, N.I.S.T. SRM 909b and Hi Chem Align linearity standards/controls. *Some control shifts should be expected!* PVI recommends that an independent correlation study be performed to confirm the appropriate operational parameters for your laboratory before utilizing this product in compliance with good laboratory practices. See note 1 for further information. DO NOT run this electrode without quad ring seal installed or replace quad ring with O Ring - leakage will occur! DO NOT USE this electrode as a Chloride electrode!

Initial Installation: This electrode has been prepared for installation at the factory. Remove the black rubber plugs and install per instrument manufacturer's instructions (these will vary slightly depending on instrument model).

Maintenance: Follow OEM recommended procedure(s) in instrument operators manual. Procedure will vary depending on the specific analyzer model.

References and Interferences:

Note 1 - A comparison of Hitachi and Performax Brand Ion-selective Electrode Response; D. Nagahara, D. Otten; *Internal White Paper*; Peripheral Visions, Inc.

For interferences see DATASHEET - Cartridge Chloride Electrode date of issue - 04122005; Peripheral Visions, inc.

**TO REORDER THIS PART PLEASE CALL 253.735.3910 or CALL
(IN U.S. TOLL FREE) 800.728.4146.**